

Sensirion AG Laubisruetistrasse 50 8712 Staefa, Switzerland

REACH- and RoHS - Declaration

We declare that all these Mass Flow Meters for Gases supplied by Sensirion AG comply with

Mass Flow Meters for Gases	REACH	RoHS
ASF14xx	yes	yes
SFM30xx	yes	yes
SFM3100, SFM3119	yes	yes
SFM3200	yes	yes
SFM3300, SFM3304, SFM3400	yes	yes
SFM4100, SFM4200, SFM4300	yes	yes
SFM5xxx	yes	yes
SFM6xxx	yes	yes
SFM70xx	yes	yes
SFS70xx	yes	yes
SGM6xxx	yes	yes
SGM70xx	yes	yes
SGM8xxx	yes	yes
Eval-Kits EK-F3x, EK-F4x, EK-F3x-CAP	yes	yes

- the EU regulation concerning the Registration, Evaluation, Authorisation and Restriction of Chemicals, REACH (Reg. 1907/2006 /EC),
 - i.e. the following limit values are not exceeded
 - Substances of Very High Concerns (SVHC)

1000 ppm

The information is based on data provided by our suppliers.

No attempt has been made to verify this information by third party analysis.

- the EU regulation on the Restriction of Hazardous Substances in electrical and electronic equipment, **RoHS** (Reg. 2011/65/EU and the delegated directive 2015/863/EU) i.e. the following limit values are not exceeded

- Lead (Pb), Mercury (Hg), Hexavalent Chromium (CrVI)

1000 ppm

- Cadmium (Cd)

100 ppm

- Polybrominated biphenyls (PBB) and Polybrominated diphenyl ethers (PBDE)

1000 ppm

- Phthalates: DEHP, BBP, DBP and DIBP

1000 ppm

- Copper alloy in the plugs of the eval-kits:

Compliant with exception 6c of 2011/65/EU.

Disclaimer: All information in this declaration is given to the best of our present knowledge and belief. The information given doesn't imply giving a warranty within the meaning of the warranty law.

Signed for and on behalf of Sensirion AG

Staefa, 14 February 2024

Dr. Adrian Ruf **EHS Manager**